

IMPORTANT NOTICE

Catalogue No's.
IA2638, IA2639, IA2640, IA2633, IA3109, IA3110, IA3111 & IA3112

Radox recommends that ACTH is to be assayed immediately
after the 30-minute reconstitution procedure..

Authorised by: Stephen Anderson
Technical Support
Date: 31st July 2017

Ref: 98/21585