Aesculap AG Quality Management

Postfach 40 78501 Tuttlingen Germany

Contact:	Christian Strobel	
Fon: Fax: Email:	+49 7461 95-31926 +49 7461 95-1555 vigilance_aag.de@aesculap.de	
Internet:	http://www.bbraun.com	
Date:	October 24, 2017	

Safety Notice - Product Recall

NR485K - COLUMBUS REV FEMUR SPACER DISTAL F5 15MM - BATCH 51468773 | 51577362 NR486K - COLUMBUS REV FEMUR SPACER DISTAL F6 15MM - BATCH 51447588 | 51503765 | 51586632 NR487K - COLUMBUS REV FEMUR SPACER DISTAL F7 15MM - BATCH 51447589 | 51468302 NR585K - COLUMBUS REV FEMUR SPACER POST.F5 15MM - BATCH 51447595 | 51503760 | 51589205 NR586K - COLUMBUS REV FEMUR SPACER POST.F6 15MM - BATCH 51447596 | 51571394 | 51586634 NR587K - COLUMBUS REV FEMUR SPACER POST.F7 15MM - BATCH 51447597 | 51585136

As part of the continuous product improvement, the dimensions of the above articles have been modified as follows.

The following figure shows the difference between the initial and the modified version of the component.




Figure 1: Difference between the initial and the modified version of the Columbus® Revision implant

The modified version of the articles is not compatible with the components of the previous version of the **Columbus® Revision System**.

According to our research, your facility has received affected articles from the Columbus[®] Revision System.

Chairman of Supervisory Board: Prof. Dr. h.c. Ludwig Georg Braun Executive Board: Dr. Joachim Schulz (Chairman) Dr. Jens von Lackum Corporate Office: Tuttlingen Register Court: Stuttgart HRB 726261 VAT reg. no. DE812160059

WEEE-Reg.-No. DE 65109852

 Bank Account:

 Deutsche Bank AG Tuttlingen

 BLZ 653 700 75 Konto 21 22 000 00

 IBAN DE44 6537 0075 0212 2000 00

 SWIFT / BIC DEUTDESS653

 Baden-Württembergische Bank

 BLZ 600 501 01 Konto 487 1905

 IBAN DE31 6005 010 0004 8719 05

 SWIFT / BIC SOLADEST

Address: Aesculap AG Am Aesculap-Platz 78532 Tuttlingen Germany Page 2 to the letter of October 24, 2017

The joint application of both versions can result in intraoperative removal of more bone material than necessary. As a result, more bone cement has to be used than planned, which can lead to a early loosening of the prosthesis. To date, we have not received any market feedback on such incident.

Our investigations have shown that the articles which are affected, can be limited to the above mentioned batches.

The affected implants can clearly be identified by comparing the batch numbers.

Please ensure that the affected implant components are no more used.

Should you have an affected product, please return it with the attached "Product Recall Form" to

Aesculap AG LRP Siegfried Schwarz Am Aesculap–Platz D-78532 Tuttlingen vigilance_aag.de@aesculap.de

For any product-related request, kindly do not hesitate to contact our product manager:

Denis Hoeffgen **2** + 49 7461 95 1785 **2** + 49 151 12635913 <u>denis.hoeffgen@aesculap.de</u>

In the case you do not have any of the affected products, please send us the attached "Feedback Form" and tick as appropriate.

Please ensure in your organization that all users of the affected devices are informed about this safety information. If you have distributed the products to a third party, please forward a copy of this information or inform the above mentioned contact person. The Competent Authority BfArM - Bundesinstitut für Arzneimittel und Medizinprodukte, has received a copy of this safety information.

We apologize for any inconvenience this may cause and thank you very much for your support.

With best regards,

Aesculap AG

i.A.

i.A.

Kerstin Rothweiler Team Leader Quality Management Vigilance Dpt. Safety Officer Medical Device Miriam Hoßfeld Quality Management Vigilance Vigilance Manager Page 3 to the letter of October 24, 2017

FEEDBACK FORM / FSCA

NR485K - COLUMBUS REV FEMUR SPACER DISTAL F5 15MM - BATCH 51468773 | 51577362 NR486K - COLUMBUS REV FEMUR SPACER DISTAL F6 15MM - BATCH 51447588 | 51503765 | 51586632 NR487K - COLUMBUS REV FEMUR SPACER DISTAL F7 15MM - BATCH 51447589 | 51468302 NR585K - COLUMBUS REV FEMUR SPACER POST.F5 15MM - BATCH 51447595 | 51503760 | 51589205 NR586K - COLUMBUS REV FEMUR SPACER POST.F6 15MM - BATCH 51447596 | 51571394 | 51586634 NR587K - COLUMBUS REV FEMUR SPACER POST.F7 15MM - BATCH 51447597 | 51585136

Please sent back this feedback form via fax or e-mail to:

Department QMV

Fax +49 7461-95 1555

vigilance_aag.de@aesculap.de

We have no affected products.

We will return the affected products.

The affected products were successfully implanted.

NAME	DEPARTMENT	_ PHONE

HOSPITAL _____ LOCATION _____

SIGNATURE _____

PRODUCT RECALL



Hygienic condition:

new good

used decontaminated

used not decontaminated

pos.	part no.	serial /	quantity	remark		
no.	article no.	lot-no.				

RETURN ADRESS :

Aesculap AG LRP Siegfried Schwarz Am Aesculap-Platz D-78532 Tuttlingen – Germany

ADRESS / SENDER:		
DATE / SIGNATURE :		

Chairman of Supervisory Board: Executive Board: Prof. Dr. h.c. Ludwig Georg Braun Dr. Joachim Schulz (Chairman) Dr. Jens von Lackum WEEE-Reg.-No. DE 65109852

Corporate Office: Tuttlingen Register Court: Stuttgart HRB 726261 VAT reg. no. DE812160059

Bank Account: Address: Aesculap AG Deutsche Bank AG Tuttlingen BLZ 653 700 75 Konto 21 22 000 00 Am Aesculap-Platz IBAN DE44 6537 0075 0212 2000 00 78532 Tuttlingen SWIFT / BIC DEUTDESS653 Germany Baden-Württembergische Bank BLZ 600 501 01 Konto 487 1905 IBAN DE31 6005 0101 0004 8719 05

SWIFT / BIC SOLADEST