



COOK MEDICAL EUROPE LTD.  
O'HALLORAN ROAD  
NATIONAL TECHNOLOGY PARK  
LIMERICK, V94 N8X2, IRELAND  
TEL: +353 61 334440 FAX: +353 61 334441  
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0011

Date: 08 October 2021

## **Urgent Field Safety Notice – Medical Device Recall** **Transseptal Needle & Transseptal Needle with Catheter**

For Attention of: Chief Executive / Risk Management / Purchasing

| Contact details of local representative (name, e-mail, telephone, address etc.)   |
|---|
| <p>Cook Medical Europe Ltd.<br/>O'Halloran Road<br/>National Technology Park<br/>Limerick, Ireland<br/>E-mail: <a href="mailto:European.FieldAction@CookMedical.com">European.FieldAction@CookMedical.com</a><br/>Phone: Please refer to the attached Country Contacts List</p> <p>For any further information or support concerning the information within this FSN, please contact your local Cook Medical Sales Representative or Cook Medical Europe Ltd.</p> |



COOK MEDICAL EUROPE LTD.  
O'HALLORAN ROAD  
NATIONAL TECHNOLOGY PARK  
LIMERICK, V94 N8X2, IRELAND  
TEL: +353 61 334440 FAX: +353 61 334441  
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0011

## **Urgent Field Safety Notice – Medical Device Recall**

### **Transseptal Needle & Transseptal Needle with Catheter**

#### **Risk Addressed by FSN**


| <b>1. Information on Affected Devices</b>                  |  |   |   |                                     |
|--|--|---|---|-------------------------------------|
| 1.   | <b>1. Device Type(s)</b><br>The products are sterile, single-use devices. The Transseptal Needle (TSNC-18-71.0 & TSNC-19-56.0) consists of a needle and obturator. The Transseptal Needle with Catheter (TSN-17-75.0-ENDRYS) is a coaxial set consisting of an outer catheter, curved-tip metal cannula, and a tapered-tip inner needle.   |   |   |                                     |
| 1.   | <b>2. Commercial name(s)</b>   | <b>3. Primary clinical purpose of device(s)</b>   | <b>4. Device Model / Catalogue / Part Number(s)</b> | <b>5. Affected lot number range</b> |
|  | Transseptal Needle   | Intended for transseptal left heart access in both diagnostic and interventional procedures | TSNC-18-71.0  | All lots                            |
|  |  |   | TSNC-19-56.0  |                                     |
|  | Transseptal Needle with Catheter   | Intended to facilitate transseptal entry into the left atrium                               | TSN-17-75.0-ENDRYS                                  |                                     |
| <b>2. Reason for Field Safety Corrective Action (FSCA)</b> |  |   |   |                                     |
| 2.   | <b>1. Description of the product problem</b><br>Cook Medical has identified that the transseptal needles may contain rust on the interior and/or exterior of the needle.   |   |   |                                     |
| 2.   | <b>2. Hazard giving rise to the FSCA</b><br>Potential adverse events that may occur if an affected product is used include increased procedural time (to obtain a replacement device) and inflammatory reactions ranging from local / self-limited reactions to systemic reactions requiring medical intervention. Systemic reactions could potentially lead to permanent impairment or be life-threatening.<br><br>To date, Cook has received no complaints reporting adverse patient effects. Cook has received four complaints where the presence of rust was identified prior to patient contact. However, please be advised that the presence of rust may go undetected by the user.  |   |   |                                     |
| <b>3. Type of Action to Mitigate the Risk</b>              |  |   |   |                                     |
| 3.   | <b>1. Action To Be Taken by the User</b><br><input checked="" type="checkbox"/> Identify Device<br><input checked="" type="checkbox"/> Quarantine Device<br><input checked="" type="checkbox"/> Return Device<br><input checked="" type="checkbox"/> Other<br>Please complete the enclosed Customer Reply Form. Where product is indicated as being returned, our Customer Services department will contact you to organize the return and issue you with the relevant Returns Authorization number. Please include contact details on the Customer Reply form.<br><br>Returned Product should be addressed to:<br>Cook Medical EUDC<br>Robert-Koch-Straße, 2<br>52499 Baesweiler<br>GERMANY<br><br>Credit will be provided for the returned affected products where applicable. |   |   |                                     |



COOK MEDICAL EUROPE LTD.  
O'HALLORAN ROAD  
NATIONAL TECHNOLOGY PARK  
LIMERICK, V94 N8X2, IRELAND  
TEL: +353 61 334440 FAX: +353 61 334441  
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0011

| 3. Type of Action to Mitigate the Risk Continued |   |     |
|--|---|-----|
| 3.   | <b>2. Is Customer Reply Required?</b><br>Form is attached specifying deadline for return.               | Yes |
| 3.   | <b>3. Action Being Taken by the Manufacturer</b><br><input checked="" type="checkbox"/> Product Removal |     |

| 4. General Information |  |  |
|------------------------|--|--|
| 4.                     | 1. FSN Type  | New  |
| 4.                     | 2. Further advice or information already expected in follow-up FSN?  | No   |
| 4.                     | 3. Manufacturer information<br>For contact details of local representative refer to page 1 of this FSN           |  |
|                        | a. Company Name  | Cook Incorporated  |
|                        | b. Address   | 750 Daniels Way<br>Bloomington, IN 47402, United States                              |
| 4.                     | 4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. |  |
| 4.                     | 5. Name/Signature  |  |
|                        |  | Larry D. Pool<br>Director, Post Market<br>Cook Incorporated                          |

| Transmission of this Field Safety Notice   |  |
|--|--|
| <p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred.</p> <p>Please transfer this notice to other organisations on which this action has an impact.</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p> |  |