



Erbe Elektromedizin GmbH P. O. Box 14 20 72004 Tuebingen Germany

SIA Arbor Medical Korporacija
Attn. Dace Ratfeldere, Dita Cerina
Augstrozes Street 2C
1026 Riga
Latvia

February 12th, 2026

FIELD SAFETY NOTIFICATION

Field Safety Corrective Action of Erbe Flexible Cryoprobes (reference number FSCA-2026.001)

Part Number	Part Description	UDI-DI	Lot Numbers
20402-401	Flexible Cryoprobe (OD 1.1mm, L1.15mm) w/ oversheath (OD2.6mm, L817mm)	04050147021785	Reference Attachment 1
20402-402	Flexible Cryoprobe (OD 1.1mm, L1.15mm) w/ oversheath (OD2.6mm, L757mm)	04050147021808	
20402-410	Flexible Cryoprobe OD 1.7mm, L1.15mm	04050147021822	
20402-411	Flexible Cryoprobe (OD 2.4mm, L1.15mm	04050147021846	

Dear valued customer,

The purpose of this letter is to advise you that Erbe is voluntarily recalling (removing) affected Erbe Flexible Cryoprobes. The Erbe Flexible Cryoprobes are intended for palliative devitalization (destruction) of tissue during interventional procedures by the application of extreme cold and cryoadhesion for applications in pulmonology such as the removal of foreign bodies, mucus plugs, blood clots, necrotic tissue, tissue tumors (palliative recanalization) and tissue biopsies.

Serious injuries have occurred or could occur due to the failure mode associated with this removal action. This rupture has only been reported on the outer white tube, the part of the probe which is outside of the patient.

Erbe Elektromedizin GmbH Waldhoernlestrasse 17 72072 Tuebingen Germany

Phone +49 7071 755-0
info@erbegroup.com
erbegroup.com

CEO and President: Christian O. Erbe
Commercial Register Court Stuttgart
HRB 380137
VAT ID (USt-IdNr.): DE 146893422

Baden-Württembergische Bank
IBAN: DE27 6005 0101 7477 5011 35
BIC: SOLADEST600

Kreissparkasse Tübingen
IBAN: DE54 6415 0020 0000 0001 41
BIC: SOLADES1TUB

Reason for Action:

Through complaints, Erbe has been made aware of cases with the affected Cryoprobes rupturing/bursting during activation. The rupture causes a loud sound which could cause tinnitus, temporary hearing loss, or persistent hearing loss. In addition, if the probe is held at the rupture/burst location, minor injuries to the hand and/or fingers could occur. The rupture/burst is caused by an insufficient adhesive application in production, leading to excessive input pressure. After a thorough investigation, it has been estimated that the reported failure mode is less than 0.1% of Cryoprobes manufactured during the identified timeframe. At the time of this letter, Erbe received 43 complaints related to this product problem. The 43 complaints have been identified as adverse events. The concerned lot numbers consist of a total quantity of 90,034 units.

Risk to Health:

The rupturing/bursting of the Cryoprobe can affect patients, healthcare professionals, or any person near the device when activated. The device failure produces a loud noise that can cause tinnitus, temporary hearing loss, or persistent hearing loss. Other potential injuries could include physical injuries from the burst, such as hand injuries and potential minor burns.

- In most reported cases, tinnitus and/or temporary hearing loss were documented.
- In one reported case, persistent hearing loss was documented.
- In three reported cases, minor injuries (i.e., hyperextended fingers, hand hurting) were documented.

Action required by recipients:

Our records indicate that you purchased one or more of the affected products. Therefore, Erbe requires that you perform the following actions:

1. Examine your inventory and quarantine any identified devices with the affected lot numbers from Attachment 1.
2. If you have further distributed these products, identify your customers/locations, and forward the notification accordingly. Please make sure to include your organization's return delivery address and contact data to facilitate the return process in your market.
3. Ensure that all personnel within the vicinity of the device activation carefully read the content of this notification.
4. Collect and quarantine the returned products in your facility. Keep them separate from another inventory. Make them available for a return delivery to the manufacturer.
5. Further operational instructions about handling the received and quarantined goods, documentation, and commercial implications for distribution partners are laid out in Attachment 3, including information about how you should create your local customer letters and the feedback form which you can use to facilitate the return process between an account in your market area and your distributing facility.
6. Within 4 weeks we anticipate the collection process to be finished. Then please fill out and send your own feedback form (Attachment 2 to this letter) based on the received overall quantities and lot numbers to us.
7. Based on your returned feedback form, we will get in contact with you and will consult on how to handle the return to the manufacturer and initiate your compensation/credit for the returned goods.

Actions taken by Erbe:

Erbe has improved and stabilized the manufacturing process and implemented additional inspection controls to ensure product specifications are met.

Please accept our sincere apologies for the inconvenience this quality measure may cause. If you have any questions or require further information, please contact our customer hotline at telephone number +49 7071 755-437, fax +49 7071 755-5437 or e-mail at techservice@erbegroup.com.

We apologize for any inconvenience caused and seek your understanding in this matter.

Yours faithfully,

Erbe Elektromedizin GmbH



i.V. Martin Viesel
Director Quality Assurance

Enclosures:

- Attachment 1: List of affected lot numbers
- Attachment 2: Feedback form, SSU/Distributor to Erbe HQ
- Attachment 3: Step by step instructions
- Attachment 4: Draft, Customer information document

Attachment 1 to the FIELD SAFETY NOTIFICATION FSCA-2026.001

LIST OF AFFECTED LOT NUMBERS

20402-401			
W4468495	WO462846	WO464700	WO468625
WO461829	WO462847	WO465022	WO468626
WO461831	WO462848	WO465023	WO468627
WO461832	WO463314	WO465054	WO468637
WO461833	WO463316	WO465055	WO468638
WO461834	WO463317	WO465056	WO468639
WO461835	WO463395	WO465057	WO468640
WO461843	WO463396	WO465247	WO468641
WO462117	WO463397	WO465248	WO468662
WO462121	WO463398	WO465249	WO468664
WO462122	WO463399	WO465266	WO468665
WO462123	WO463407	WO465267	WO468669
WO462374	WO463408	WO465268	WO469358
WO462387	WO463409	WO465269	WO469359
WO462388	WO464270	WO467858	WO471335
WO462389	WO464278	WO467886	WO471336
WO462831	WO464279	WO468210	WO471337
WO462838	WO464699	WO468612	WO471338
			WO471339

20402-402			
WO461307	WO461313	WO461830	WO463315

20402-410			
W4465597	WO463323	WO464284	WO468643
WO462187	WO463324	WO465059	WO468644
WO462322	WO463325	WO465060	WO468645
WO462390	WO463326	WO465061	WO468647
WO462391	WO463386	WO465062	WO468671
WO462392	WO463387	WO467876	WO468672
WO462393	WO464280	WO467877	WO468673
WO462396	WO464281	WO467878	WO468674
WO462677	WO464282	WO467879	WO469361
WO462678	WO464283	WO468225	WO469362
			WO469363

20402-411			
W4465351	WO463327	WO464286	WO467874
WO461840	WO463328	WO465063	WO467875
WO461846	WO463388	WO465064	WO468648
WO462097	WO464285	WO467873	WO468649
			WO469364